

Calibration certificate

Part No. 500001/500002



Calibration Certificate

The calibration is carried out over a measurement range from 0.4 m to 16 m under stable ambient conditions. The reference device refers to national and international standards.

Calibration Certificate for new Laser Distance Sensor

For new Laser Sensors, a calibration certificate can be ordered as an optional item at the time of purchase by specifying the corresponding separate part number (Part No. 500001).

Recalibration for used Laser Distance Sensor

For used Laser Sensors returned for recalibration, the sensor and its lens will be thoroughly cleaned to ensure precise measurement capability. The sensor will be calibrated over the specified measurement range under stable environmental conditions.

Calibration Certificate

The calibration certificate shows the current measurement values in graphical and tabular form.

Any necessary repairs are not included in the recalibration service.

Order information

Part No.	Description
500001	Calibration certificate for new Laser Sensor measuring devices (must be ordered together with the device)
500002	Recalibration with calibration certificate for used Laser Sensor measuring devices

